



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT



RECEIVED
Oriss Breath Alcohol Program
By Carol Day at 10:38 am, Sep 08, 2009

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN <u>950064 - Pettis Co. Jail</u>	DATE OF INSPECTION <u>08-29-2009</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>319 S. Lamine Ave., Sedalia, MO 65301</u>	TIME OF INSPECTION <u>1249 hours</u>

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	
<input type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> TIME AND DATE	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34 °C ± 0.2°C) <u>34.0°C</u>	

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

- ☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE
 (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 <input checked="" type="checkbox"/> .102	TEST 2 <input checked="" type="checkbox"/> .102	TEST 3 <input checked="" type="checkbox"/> .103
---	---	---

<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)					
<input checked="" type="checkbox"/> NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)					
REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>1</u>	(Over .19) <u>0</u>

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

This instrument conforms to the Department of Health standards.

Supplier: Replo Concentration: .100% Exp Date: 10-13-2010
 Lot #: 08002

SPECTING OFFICER

SIGNATURE <u>Victoria A. Kottman</u>	PRINT NAME <u>Victoria A. Kottman</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>720249</u> Exp. Date: <u>12-18-2009</u>	TELEPHONE NUMBER <u>660/826-8100</u>

REPCO MARKETING INC.

3101-155 STONYBROOK DRIVE
RALEIGH, N.C. 27601
919-278-5260

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 08002
EXPIRATION DATE: October 13, 2010 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 08002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed utilizing a gas chromatograph and found to contain 0.1209 gms/dl wt/vol ethyl alcohol in aqueous solution (i.e. ethanol).

The alcohol and distilled water used in the solution were found to be free of any interfering substances.

This solution will produce a vapor alcohol value of 0.100% +/-3% when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator.

The date of manufacture for this lot number is October 14, 2008. The expiration date for this lot number is October 13, 2010 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Cecil E. Gerner, President
RepCo Marketing, Inc.

BAC DataMaster Evidence Ticket

REF ID: A66763 SERIAL NUMBER 00000004

— 財政部 —

PROGRAM 054-87-2009: DKRY

THE UNIVERSITY OF CHICAGO

THE
BIBLICAL
ARCHAEOLOGIST

SECRET

2014年12月

「それは、さういふことではなからうか。それならば、さういふことは、さういふことではなからうか。」

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SERIALS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 952254
08/29/09

TESTING OFFICER:
KOTTMAN/W/R
OFFICER I.D. #: 170
PERMIT NUMBER: 720245
EXPIRATION DATE: 12/18/09
MISCELLANEOUS DATA:

--- SUPERVISION MODE ---

BLANK TEST	.202	12:02
INTERNAL STANDARD	VERIFIED	12:02
EXTERNAL STANDARD	.102	12:02
BLANK TEST	.000	12:03
EXTERNAL STANDARD	.102	12:03
BLANK TEST	.000	12:04
EXTERNAL STANDARD	.103	12:04
BLANK TEST	.000	12:05

N = 3
SIM. = .1
AVG. = .1223

Operator Signature

Victoria Kottman #170

Face This Side Down - This Edge In First

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
SEDALIA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 952224
02/09/09

ARREST TIME: 12:00
SUBJECT NAME:

TEST 1/8

DOB: 12/12/44

SEX: F

STATE/D.L.I. NO/4

ARRESTING OFFICER:

KOTTMAN/V/R

OFFICER I.D.# 170

TESTING OFFICER:

KOTTMAN/V/R

OFFICER I.D.# 170

PERMIT NUMBER: 726249

EXPIRATION DATE: 12/12/09

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	000	12:00
INTERNAL STANDARD	VERIFIED	12:00
RADIO INTERFERENCE		

Operator Signature

Victor A. Kottman #170

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



VICTORIA KOTTMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER


for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/18/07

Number 720249

Expires 12/18/2009

MO 580-0771 (7-88)


Director of State Public Health Laboratory


Director, Department of Health

Lab. 4 (R7-88)